



Medline Industries Ltd
3rd Floor, Quayside Wilderspool Business
Park
Greenalls Avenue
Warrington WA4 6HL
United Kingdom

Tel.: +44 844 334 5237
Fax: +44 844 334 5238
www.medline.eu/uk



Technical Sheet

ENDOSID[®]

Endotracheal Tube, sterile

Reference(s)

Reference	Pack factor (Box)	Pack Factor (Case)	Ø Interior Diameter (Mm)	Ø Exterior Diameter (Mm)	Tubing Length in mm
841030A	10	100	3	4.2	165
841035A	10	100	3.5	4.9	185
841040A	10	100	4	5.5	210
841045A	10	100	4.5	6.2	225
841050A	10	100	5	6.9	245
841055A	10	100	5.5	7.5	275
841060A	10	100	6	8.2	285
841065A	10	100	6.5	8.8	295
841070A	10	100	7	9.6	310
841075A	10	100	7.5	10.2	315
841080A	10	100	8	10.9	330
841085A	10	100	8.5	11.5	330
841090A	10	100	9	12.1	330
841095A	10	100	9.5	12.7	330
841100A	10	100	10	13.6	330

Pre-curved endotracheal tube (Magill) with cuff and Murphy eye for securing the airway in the field of anesthesia and emergency medicine in oral and nasal use.
Universal connector at the upper end for connection to a ventilator or resuscitator bag.

It should not be used:

- in patients who are sensitive to any of the materials used
- In combination with oil-based lubricants

Not suitable for Use in MRI

- Shelf Life 5 years
- Storage +15°C to +49°C, in a dry place, away from sunlight
- Single-Use product, not reusable
- Not reprocessible
- Universal connector 15 mm
- Consistent X-ray contrast stripe
- Blue pilot balloon with valve for Luer and Luer-Lock syringes
- Double black marking ring to control the depth of intubation

Description



0123 Latex free

DEHP free

• Length graduation

Main materials

Product part	Material
Tube	Polyvinyl chloride, clear (PVC)
Cuff	Polyvinyl chloride (PVC)
Universal connector	Polypropylene (PP)
Inflation balloon	Polyvinyl chloride (PVC)
Valve inflation balloon	RPVC, PP, ABS, stainless steel
X-ray contrast strip	Barium sulphate (BaSo4)

Legal Manufacturer

Well Lead Medical Co., Ltd.
C-4 Jinhu Industrial Estate, Hualong 511434 Panyu, Guangzhou
P.R. CHINA
In compliance with ISO13485

Regulation

In compliance with Medical Devices Regulations 2017/745
MD Class IIa- Sterile

•Please refer to complete labeling before using the product.

