

## 657200 Medical Face Mask Standard

### Product details

---

**Descriptive feature:** Ear Loop, Type II

**Color:** Blue

**Sterility:** Non-sterile

### Images

---



### Delivered items

---

#### 657200-10

**Country of origin:** France

**Shelf life:** 5 years

**Sterilization method:** Non-sterile

**Packing information:** First packaging layer is a paper board dispenser box. Second layer is a corrugated board transport box.

Packaging level	Pack count	GS1 code
Dispenser box	50	7310791114217
Transport box	600	7310791114224
Pallet	33600	7332430502063

## Material

---

### Animal tissues:

No

### Human blood derivatives:

No

### Medicinal substances:

No

### Phthalates:

No

### Polyvinyl chloride:

No

## Product Composition Medical Face Masks, ear-loop

Product Component	Composition
Outer facing	Polypropylene nonwoven
Inner facing	Polypropylene nonwoven
Filter medium	Polypropylene nonwoven
Nose clip	Polypropylene covered steel
Ear loops	Polypropylene / Polyurethane

## Product Performance Medical Face mask

Characteristics	Test Method	Internal Test Method	Unit	Requirement	Product Performance
Bacterial filtration efficiency (BFE)	EN 14683, App B	N/A	%	≥98	Pass
Breathability	EN 14683, App C	N/A	Pa/cm <sup>2</sup>	<29.4	Pass
Splash resistance	ISO 22609	N/A	kPa	Not specified	N/A
Microbial cleanliness	ISO 11737-1	T-303	CFU/g	≤30	Pass

## Technical

---

## Dimension

Dimension text	Dimension value
Length	175 mm
Width unfolded	180 mm
Width folded	85 mm
Length ear-loop (unstretched)	155 mm
Nose clip	125 mm

## Disposal instructions

Non-hazardous waste used BARRIER products and sterility barriers should, in the majority of cases, be classified as non-hazardous waste. They contain high amounts of energy and are well suited for incineration. BARRIER products do not contain any hazardous substances that can leach out if the products are land filled. Transport boxes are designed to fit existing recovery systems. The new BARRIER packaging system complies with the Packaging Waste Directive of the European Union.

## Storage instructions

Mölnlycke Health Care recommends that BARRIER products are stored under normal storage conditions. All layers of packaging should be kept intact until access to the underlying layers is needed. Storage facilities for products only protected by the sterility barrier should be kept under conditions where low level of particulate air contamination prevail, so that it would not constitute a risk to the patient when the package is opened and the product is used.

## Classification

<b>Regulation type</b>	<b>MDD Class I</b>
<b>MDD Classification Rule:</b>	1
<b>Intended use MDD:</b>	The device is intended to be worn to protect both the patient and healthcare personnel from the transfer of microorganisms, body fluids and particulate matter.

**Applied standards:** The standards presented below is a selection of the most essential standards that are adhered to.

EN 1041, EN 14683, EN 62366, EN ISO 9001, EN ISO 13485, EN ISO 10993-1, EN ISO 10993-5, EN ISO 15223-1, ISO 15223-2, EN ISO 10993-10, ISO 14001

## Removable Label

No

## GMDN Code (Global Medical Device Nomenclature)

35177

## Reusability

Single use